



June 11, 2015

K&R TECHNOLOGIES LLC
MICROBIOLOGY INTERNATIONAL
CHRISTOPHER HAAG
5111 PEGASUS COURT, SUITE H
FREDERICK MD 21704

SYSTEC GMBH LABOR - SYSTEMTECHNIK
RALF RICKER
SANDUSWEG 11 D-35435
GERMANY

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: SYSTEC GMBH LABOR - SYSTEMTECHNIK
Product Name: SYSTEC AUOTCLAVES (trans id 2553099)
Model Number(s): DB23, DB45, DB65, DB90, DB100, DB150, DB200, DB250, DE23, DE45, DE65, DE90, DE100, DE150, DE200, DE250, DX23, DX45, DX65, DX90, DX100, DX150, DX200, DX250, HX65, HX90, HX100, HX150, HX200, HX210, HX320, HX430, HX540, HX580, HX650, HX780, HX980, HX1180 HX1580, HX 2 DOOR 90, HX 2 DOOR 150, HX 2 DOOR 200, HX 2 DOOR 320, HX 2 DOOR 430, HX 2 DOOR 540, HX 2 DOOR 580, HX 2 DOR 650, HX 2 DOOR 780, HX 2 DOR 980, HX 2 DOOR 1180, HX 2 DOOR 1580, MP10, MP20, MP30, MP45, MP65, MP90, MP120, VB40, VB55, VB65, VB75, VB95, VB100, VB120, VB150, VE40, VE55, VE65, VE75, VE95, VE100, VE120, VE150, VX40, VX55, VX65, VX75, VX95, VX100, VX120 AND VX150.
Product File No: 20150150

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of June 2020.

This approval is contingent upon compliance with the following stipulation(s):

A reduced pressure principle backflow preventer (RPZ) that conforms to ASSE 1013 or CAN/CSA B64.4 shall be installed in each fixture supply serving this product. In addition, the RPZ shall:

1. be installed in accordance with s. SPS 382.41 (5) (f) Wis. Adm. Code;
2. be tested in accordance with s. SPS 382.22 (8); and
3. be registered with this department in accordance with s. SPS 382.20 (1) (c).

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail